

Chapter 12 Noise

Appendix 12.1 Noise Monitoring Locations

Appendix 12.1: Noise Monitoring Locations



Figure B.1: Sound level meter on a tripod on Bradford Road



Figure B.2: Sound level meter on a tripod on Cliff Hollins Lane



Figure B.3: Sound level meter on a tripod on west boundary of residential zone, along Cliff Hollins Lane

Appendix 12.2 Calibration Certificates

Campbell Associates Ltd
 5b Chelmsford Road Industrial Estate
 GREAT DUNMOW, CM6 1HD, England
www.campbell-associates.co.uk
info@campbell-associates.co.uk
 Phone 01371 871030 Facsimile 01371879108



Certificate number: U13982

Certificate of Calibration and Conformance

Test object: Sound Calibrator
Manufacturer: Brüel and Kjær
Type: 4231
Serial no.: 2328978
Customer: AECOM
Address: 6th Floor, One New York Street,
 Manchester, M1 4HD
Contact Person: Jon Casey

Measurement Results	Level	Level Stability	Frequency	Frequency Stability	Distortion
1.	93.91 dB	0.01 dB	999.97 Hz	0.00 %	0.35 %
2.	93.91 dB	0.01 dB	999.97 Hz	0.00 %	0.35 %
3.	93.91 dB	0.01 dB	999.97 Hz	0.00 %	0.35 %
Result (Average):	93.91 dB	0.01 dB	999.97 Hz	0.00 %	0.35 %
Expanded Uncertainty	0.10 dB	0.02 dB	1.00 Hz	0.01 %	0.10 %
Degree of Freedom	>100	>100	>100	>100	>100
Coverage Factor	2.00	2.00	2.00	2.00	2.00

The stated level is relative to 20µPa. The level is traceable to National Standards.
 The stated level is valid at reference conditions. The following correction factors have been applied during the measurement: Pressure: 0.00008 dB/kPa Temperature: 0.0015 dB/°C Relative humidity: 0.001 dB%/RH Load volume: 0.0003 dB/mm³
 The reported expanded uncertainty of measurements is based on a standard uncertainty multiplied by the coverage factor of k=2, providing a level of confidence of approximately 95%. Where the degrees of freedom are insufficient to maintain this confidence level, the coverage factor is increased to maintain the confidence level. The uncertainty has been determined in accordance with UKAS requirements.
 Records: K:\C:\Calibration\Non-ISO\Nist-1015 Cal\G031368N4231\1_2328978_M1.rtf

Environmental conditions:
 Reference conditions: Pressure: 101.325 kPa Temperature: 20.0 °C Relative humidity: 50 %RH
 Measurement conditions: 102.783 ± 0.043 kPa 22.1 ± 0.1 °C 50.9 ± 2.0 %RH

Date received for calibration: 03/07/2013
 Date of calibration: 08/07/2013
 Date of issue: 09/07/2013
 Engineer:

Supervisor

Michael Tickner

Darren Batten TechnOA

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to recognised national standards, and to the units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced otherwise in full without the written approval of the issuing laboratory.

Calibration Report

Certificate No.:14442

Norsonic Type: 118 Serial no: 31509

Customer: AECOM
Address: 6th Floor, 1 New York Street,
Manchester, M1 4HD.
Contact Person: Jon Casey.

Instrument software version: V2.0.752
Microphone: GRAS Type: 40AF Serial no: 85858 Sens: 25.86dB
Preamplifier: Norsonic Type: 1206 Serial no: 30651
Calibrator: Norsonic Type: 1251 Serial no: 27485 Level:114.06dB

Mains adapter was included
This sound level meter has been calibrated as specified in BS 7580, PART 1, 1997. The results are traceable to NPL, UK
Measurement Results:

Calibration of sound level meter - BS7580 Clause 5.4	Passed
Noise test - BS 7580 Clause 5.5.2	Passed
Level Linearity Test - BS 7580, Clause 5.5.3	Passed
Frequency weightings: A Network - BS 7580 Clause 5.5.4	Passed
Frequency weightings: C Network - BS 7580 Clause 5.5.4	Passed
Frequency weightings: Z Network - BS 7580 Clause 5.5.4	Passed
Time weightings F and S - BS7580 Clause 5.5.5	Passed
Peak response - BS7580 Clause 5.5.6	Passed
RMS accuracy - BS7580 Clause 5.5.7	Passed
Time weighting I - BS7580 Clause 5.5.8	Passed
Integrating Test: Time averaging - BS7580 Clause 5.5.9	Passed
Integrating Test: Pulse range - BS7580 Clause 5.5.10	Passed
Integrating Test: Sound exposure level - BS7580 Clause 5.5.11	Passed
Overload SPL Test - BS 7580 Clause 5.5.12	Passed
Overload Leq Test - BS 7580 Clause 5.5.12	Passed
Acoustic tests - BS 7580 Clause 5.4 and 5.8	Passed
Summation of acoustic tests - BS 7580 Clause 5.5.4	Passed

The sound level meter in the configuration tested conforms to the requirements of BS 7580 Part 1.

Comment:
Correct level with associated calibrator is 113.9dB(A)

Environmental conditions:
Pressure: 99.482 kPa
Temperature: 23.0 °C
Relative humidity: 48.5 %RH
Date of calibration: 16/09/2013
Date of issue: 16/09/2013
Supervisor: Darren Betten Tech/OA
Engineer:

Michael Tickner

Software version: 6.0d



Campbell Associates

www.campbellassociates.co.uk

Campbell Associates Ltd
 5b Chelmsford Road Industrial Estate
 GREAT DUNMOW, CM6 1HD, England
www.campbell-associates.co.uk
info@campbell-associates.co.uk
 Phone 01371 871030 Facsimile 01371879106



Certificate number: U15805

Certificate of Calibration and Conformance

Test object: Sound Calibrator
Manufacturer: Norsonic
Type: 1251
Serial no: 27485
Customer: AECOM
Address: 6th Floor, One New York Street,
 Manchester, M1 4HD.
Contact Person: Jon Casey

Measurement Results	Level	Level Stability	Frequency	Frequency Stability	Distortion
1:	114.03 dB	0.06 dB	999.40 Hz	0.00 %	<0.3 %
2:	114.02 dB	0.06 dB	999.40 Hz	0.00 %	<0.3 %
3:	114.03 dB	0.06 dB	999.40 Hz	0.00 %	<0.3 %
Result (Average):	114.03 dB	0.06 dB	999.40 Hz	0.00 %	<0.3 %
Expanded Uncertainty:	0.10 dB	0.02 dB	1.00 Hz	0.01 %	0.12 %
Degree of Freedom:	>100	>100	>100	>100	27
Coverage Factor:	2.00	2.00	2.00	2.00	2.13

The stated level is relative to 20µPa. The level is traceable to National Standards.
 The stated level is valid at reference conditions. The following correction factors have been applied during the measurement: Pressure: 0.0005 dB/kPa Temperature: 0.003 dB/°C Relative humidity: 0.000 dB/%RH Load volume: 0.0003 dB/mm³

The reported expanded uncertainty of measurements is based on a standard uncertainty multiplied by the coverage factor of k=2, providing a level of confidence of approximately 95%. Where the degrees of freedom are insufficient to maintain this confidence level, the coverage factor is increased to maintain this confidence level. The uncertainty has been determined in accordance with UKAS requirements.

Records: K:\C A\Calibration\Nor-1504\Nor-1018 CalCal\2014\NOR1251_27485_M1.rmf

Environmental conditions:	Pressure:	Temperature:	Relative humidity:
Reference conditions:	101.325 kPa	23.0 °C	50 %RH
Measurement conditions:	99.774 ± 0.043 kPa	23.0 ± 0.1 °C	48.1 ± 1.5 %RH

Date received for calibration: 11/03/2014
 Date of calibration: 20/03/2014
 Date of issue: 20/03/2014
 Engineer:

Michael Tickner

Supervisor:

Dayid Egan - Laboratory Manager

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to recognised national standards, and to the units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full without the prior written approval of the issuing laboratory.



Certificate of Calibration

Certificate No.: CAL 022-2015-5248



Test object: Sound Calibrator
Manufacturer: Norsonic
Type: 1251
Serial no: 34393

Customer:

	Level	Level Stability	Frequency	Frequency Stability	Distortion
Measurement Results:	114,01 dB	0,05 dB	1000,08 Hz	0,00 %	0,40 %
Expanded Uncertainty:	0.11 dB	0.02 dB	1.0 Hz	0.1 %	0.2 %

The stated levels are relative to 20µPa.

The stated level is valid at reference conditions. The following correction factors have been applied during the measurement:

Pressure: 0,0005 dB/kPa Temperature: 0,000 dB/°C Relative humidity: 0,000 dB/%RH Load volume : 0,0003 dB/mm³

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k, which for a t-distribution with the reported effective degree of freedom corresponds to coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA publication EA-4/02.

Records: L:\PROJECTS\CALLAB\PROGRAM\Cof2015\NOR1251_34393_M1.nmf

Environmental conditions:	Pressure:	Temperature:	Relative humidity:
Reference conditions:	101,325 kPa	23,0 °C	50 %RH
Measurement conditions:	99,000 ± 0,010 kPa	23,7 ± 0,2 °C	37,5 ± 1,0 %RH

Date received for calibration:

Date of calibration: 2015-04-09
Date of issue: 2015-04-09
Engineer: Terje Hansen
Supervisor:


Terje Hansen

This certificate of calibration is issued by a laboratory accredited by Norwegian Accreditation (NA). NA is one of the signatories to the EA Multilateral Agreement for mutual recognition of calibration certificates (European Co-operation for Accreditation). The accreditation states that the laboratory meets the NA requirements concerning competence and calibration system for all the calibrations contained in the accreditation. It also states that the laboratory has a satisfactory quality assurance system and traceability to accredited or national calibration laboratories. This certificate may not be reproduced other than in full.

Calibration Report

Certificate Number:-16528

Manufacturer: GRAS
Type: 40HL-N
Serial no: 123948


Customer: AECOM
Address: 8th Floor, One New York Street,
Manchester, M1 4HD.
Contact Person: Jon Casey.

Measurement Results:

Sensitivity:
(dB re 1V/Pa)

1:	-1.48
2:	-1.49
3:	-1.51

Result (Average): -1.49
Expanded Uncertainty: 0.10
Degree of Freedom: >100
Coverage Factor: 2.00



The following correction factors have been applied during the measurement:
Pressure:0.000 dB/kPa Temperature:0.000 dB/°C Relative humidity:0.000 dB/%RH

Reference Calibrator: WSC6 - B&K-4231-1882939 Volume correction: 0.000 dB
Records:K:\C A\Calibration\Nor-1504\Nor-1017 MicCal\2014\GRAS40HL-N_123948_M2.nmf
Measurement procedure: TP05
All results quoted are directly traceable to National Physical Laboratory, London

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2$, which for a normal distribution corresponds to coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA publication EA-4/02.

Comment:

Environmental conditions:

Pressure: 101.447 ± 0.040 kPa Temperature: 22.1 ± 0.4 °C Relative humidity: 50.1 ± 2.6 %RH

Date of calibration: 03/07/2014

Date of issue: 03/07/2014

Supervisor : David Egan, Laboratory Manager
Engineer :

Ian Campbell MSc MIOA
Software version: 6.0h

Campbell Associates
www.campbell-associates.co.uk

CERTIFICATE OF CALIBRATION

ISSUED BY AV CALIBRATION

Date of issue 08 January 2014 Certificate N^o 07487



AV Calibration
2 Warren Court
Chicksands, Shefford
Bedfordshire SG17 5QB
U.K.
Tel: +44 (0)1462 638600
Fax: +44 (0)1462 638601
Email: lab@avcalib.co.uk
www.avcalibration.co.uk

Page 1 of 3 Pages

Approved Signatory

G. Parry [] B. Baker [✓]

Acoustics Noise and Vibration Ltd trading as AV Calibration

CLIENT URS Infrastructure & Environment UK Ltd
12 Regan Way
Chetwynd Business Park
Chilwell
Notts
NG9 6RZ

F.A.O. David Gerard

ORDER No CH 73253 Job No UKAS13/12271/02

DATE OF RECEIPT 11 December 2013

PROCEDURE AV Calibration Engineer's Handbook, section 25: periodic testing of sound level meters to IEC 61672-3:2006 (BS EN 61672-3:2006) as modified by UKAS TPS 49 Edition 2: June 2009

IDENTIFICATION Sound level meter Norsonic type 140 serial No 1403909 connected via a preamplifier type 1209 serial No 13283 to a half-inch microphone type GRAS 40AF serial No 102590. Associated calibrator Brüel & Kjær type 4231 serial No 3005464 with a one-inch housing and adapter type UC 0210 for half-inch microphone.

CALIBRATED ON 08 January 2014

PREVIOUS CALIBRATION Calibrated on 20 December 2011, Certificate No. 06278 issued by this laboratory.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.